



Abtech Safety Ltd
Unit 1&2 Parkway Business Centre
Sixth Avenue
Deeside Industrial Estate
Flintshire
CH5 2LE
Tel: +44(0)1244 837 050
sales@abtechsafety.com
www.abtechsafety.com

21/06/2022

Products Supplied to Saviour Medical

ABSS – Adjustable Stretcher Lifting Bridle

Certified to EN 358:2018 under notified body 2777 (SATRA)

The test involves holding a static load of 15kN for 3 minutes without failure.

This product has been approved with a Working Load Limit of 400kg

Drag Straps and Carry Handles

Do not fall under PPE Regulations.

The 25mm Webbing has a minimum breaking strain of 20kN.

We carry out annual stitch block tests to ensure all stitch patterns take a minimum of 12kN. The latest test took the 25mm Webbing with Bar stitching to 15.61kN and 15.43kN without failure.

Technical Stretchers

Do not fall under PPE Regulations.

Internal testing has been carried out to ensure the strength of the lifting eyes.

Using bariatric dummies to distribute the weight evenly, the stretchers were tested to 440kg without failure.

A handwritten signature in black ink, appearing to read "Sally Love".

Sally Love
Business Manager